

CERTIFICATE

Survey of 13 March 2026

You have fulfilled the requirements of the External Quality Assessment with the following analysis

Autoimmune Diseases - Connective Tissue Diseases (ANA) (251):

Validity 12 months:

Collagenosis Screening (ANA) - qualitative (R: B2)

anti-DFS70 / LEDGF - qualitative

Autoantibodies against dsDNA - qualitative (R: B2)

(R) analysis is subject to the RiliBÄK

Customer:
10011570
TestLine Clinical Diagnostics s.r.o.
Krizikova 188/68
61200 Brno

Düsseldorf, 27 April 2026



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Dr. rer. nat. Martin Blüthner
(Adviser)

CERTIFICATE OF PARTICIPATION

Survey of 13 March 2026

You have participated in the External Quality Assessment with the following analysis

Autoimmune Diseases - Connective Tissue Diseases (ANA) (251):

Collagenosis Screening (ANA) - qualitative (R: B2)


anti-DFS70 / LEDGF - qualitative

Autoantibodies against dsDNA - qualitative (R: B2)

(R) analysis is subject to the RiliBÄK

Customer:
10011570
TestLine Clinical Diagnostics s.r.o.
Krizikova 188/68
61200 Brno

Düsseldorf, 27 April 2026



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Dr. rer. nat. Martin Blüthner
(Adviser)

Listing and evaluation of the results

10011570: TestLine Clinical Diagnostics s.r.o.

Survey of 13 March 2026

Adviser: **Dr. rer. nat. Martin Blüthner**
 MVZ Labor PD Dr. Volkmann und Kollegen SE & Co. eGmbH
 Gerwigstrasse 67
 76131 Karlsruhe

Tel.: +49 721 85000 225
 Fax: +49 721 85000 115
 Mail: bluetner@laborvolkmann.de

251

Autoimmune Diseases - Connective Tissue Diseases (ANA)

Collagenosis Screening (ANA) - qualitative +

Name of test	Unit	Method	Manufacturer			
		Immunoblot (247)	TEST-LINE (TT)			
Result (qual)	Sample 251A- 260224	Your specification(s)	Correct specification(s)	TV- Type		Meets criteria
	-01	positive (1)	positive (1)	SV		+
	-02	positive (1)	positive (1)	SV		+

anti-DFS70 / LEDGF - qualitative +

Name of test	Unit	Method	Manufacturer			
		Immunoblot (247)	OTHER MANUFACTURERS (ZY)			
Result (qual)	Sample 251A- 260224	Your specification(s)	Correct specification(s)	TV- Type		Meets criteria
	-01	negative (0)	negative (0)	SV		+
	-02	negative (0)	negative (0)	SV		+

Autoantibodies against dsDNA - qualitative +

Name of test	Unit	Method	Manufacturer			
		Immunoblot (247)	TEST-LINE (TT)			
Result (qual)	Sample 251A- 260224	Your specification(s)	Correct specification(s)	TV- Type		Meets criteria
	-01	negative (0)	negative (0)	SV		+
	-02	negative (0)	negative (0)	SV		+

Additional information on the evaluation is available for download in the RV-Online system in the column "Report".



Individual summary of results

10011570: TestLine Clinical Diagnostics s.r.o.

Survey of 13 March 2026

251

Autoimmune Diseases - Connective Tissue Diseases (ANA)

Collagenosis Screening (ANA) - qualitative (N = 100, Rate of success: 94%)

Sample 251A-260224-01

Collective	negative	positive	total
all methods	3	97 ●	100

Sample 251A-260224-02

Collective	negative	positive	total
all methods	4	95 ●	99

The point corresponds to the correct result, the horizontal bar corresponds to your specification, the vertical bar corresponds to your collective

anti-DFS70 / LEDGF - qualitative (N = 159, Rate of success: 79,9%)

Sample 251A-260224-01

Collective	negative	positive	total
all methods	128 ●	31	159

Sample 251A-260224-02

Collective	negative	positive	total
all methods	158 ●	1	159

The point corresponds to the correct result, the horizontal bar corresponds to your specification, the vertical bar corresponds to your collective

Autoantibodies against dsDNA - qualitative (N = 144, Rate of success: 93,8%)

Sample 251A-260224-01

Collective	negative	positive	total
all methods	136 ●	8	144

Sample 251A-260224-02

Collective	negative	positive	total
all methods	142 ●	1	143

The point corresponds to the correct result, the horizontal bar corresponds to your specification, the vertical bar corresponds to your collective